

ANALYTICAL REPORT

Job Number: 420-111923-1

SDG Number: Monticello St John St High/Middle School

Job Description: Sullivan County BOCES

For:

Sullivan County BOCES

6 Weirk Avenue

Liberty, NY 12754

Attention: Jesse Morrill



Meredith W Ruthven

Customer Service Manager

mruthven@envirotestlaboratories.com

11/17/2016

NYSDOH ELAP does not certify for all parameters. EnviroTest Laboratories does hold certification for all analytes where certification is offered by ELAP unless otherwise specified in the Certification Information section of this report. Pursuant to NELAP, this report may not be reproduced, except in full, without written approval of the laboratory. EnviroTest Laboratories Inc. certifies that the analytical results contained herein apply only to the samples tested as received by our laboratory. All questions regarding this report should be directed to the EnviroTest Customer Service Representative.

EnviroTest Laboratories, Inc. Certifications and Approvals: NYSDOH 10142, NJDEP NY015, CTDOH PH-0554

METHOD SUMMARY

Client: Sullivan County BOCES

Job Number: 420-111923-1
SDG Number: Monticello St John St High/Middle School

Description	Lab Location	Method	Preparation Method
Matrix: Water			
ICPMS Metals by 200.8	EnvTest	EPA 200.8 Rev.5.4	
200 Series Drinking Water Prep Determination Step	EnvTest		EPA 200

Lab References:

EnvTest = EnviroTest

Method References:

EPA = US Environmental Protection Agency

METHOD / ANALYST SUMMARY

Client: Sullivan County BOCES

Job Number: 420-111923-1

SDG Number: Monticello St John St High/Middle School

Method	Analyst	Analyst ID
EPA 200.8 Rev.5.4	Palentino, Gus J	GJP
EPA 200.8 Rev.5.4	Sirico, Derek	DS

SAMPLE SUMMARY

Client: Sullivan County BOCES

Job Number: 420-111923-1
SDG Number: Monticello St John St High/Middle School

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
420-111923-1	001	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-2	002	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-3	003	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-4	004	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-5	005	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-6	006	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-7	007	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-8	008	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-9	009	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-10	010	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-11	011	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-12	012	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-13	013	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-14	014	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-15	015	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-16	016	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-17	017	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-18	018	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-19	019	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-20	020	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-21	021	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-22	022	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-23	023	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-24	024	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-25	025	Drinking Water	10/18/2016 0700	10/18/2016 1545
420-111923-26	026	Drinking Water	10/18/2016 0700	10/18/2016 1545

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 001
Lab Sample ID: 420-111923-1

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1552		
Prep Method: 200			Date Prepared: 11/08/2016 1015		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 002
Lab Sample ID: 420-111923-2

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1553		
Prep Method: 200			Date Prepared: 11/08/2016 1015		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 003
Lab Sample ID: 420-111923-3

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1553	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 004
Lab Sample ID: 420-111923-4

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1554	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	6.21	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 005
Lab Sample ID: 420-111923-5

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1555	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 006
Lab Sample ID: 420-111923-6

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1556	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 007
Lab Sample ID: 420-111923-7

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1556		
Prep Method: 200			Date Prepared: 11/08/2016 1015		
Pb	16.8	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 008
Lab Sample ID: 420-111923-8

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1557	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	3.08	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 009
Lab Sample ID: 420-111923-9

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1558	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	3.61	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 010
Lab Sample ID: 420-111923-10

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1559	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	2.21	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 011
Lab Sample ID: 420-111923-11

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1601	
Prep Method: 200			Date Prepared:	11/08/2016 1015	
Pb	16.9 g	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 012
Lab Sample ID: 420-111923-12

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1453	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	2.08	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 013
Lab Sample ID: 420-111923-13

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1456	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 014
Lab Sample ID: 420-111923-14

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1456	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	6.09	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 015
Lab Sample ID: 420-111923-15

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1457	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	9.13	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 016
Lab Sample ID: 420-111923-16

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1500	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 017
Lab Sample ID: 420-111923-17

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1500		
Prep Method: 200			Date Prepared: 11/08/2016 1055		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 018
Lab Sample ID: 420-111923-18

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1501	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	1.98	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 019
Lab Sample ID: 420-111923-19

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1502	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	4.74	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 020
Lab Sample ID: 420-111923-20

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1503		
Prep Method: 200			Date Prepared: 11/08/2016 1055		
Pb	58.9 g	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 021
Lab Sample ID: 420-111923-21

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1503	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	32.4 g	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 022
Lab Sample ID: 420-111923-22

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/12/2016 1252	
Prep Method: 200			Date Prepared:	11/09/2016 1700	
Pb	7.62	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 023
Lab Sample ID: 420-111923-23

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1504	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	43.9 g	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 024
Lab Sample ID: 420-111923-24

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 11/08/2016 1505		
Prep Method: 200			Date Prepared: 11/08/2016 1055		
Pb	64.7 g	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 025
Lab Sample ID: 420-111923-25

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1506	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	5.65	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-111923-1
Sdg Number: Monticello St John St High/Middle School

Client Sample ID: 026
Lab Sample ID: 420-111923-26

Date Sampled: 10/18/2016 0700
Date Received: 10/18/2016 1545
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	11/08/2016 1506	
Prep Method: 200			Date Prepared:	11/08/2016 1055	
Pb	32.3 g	ug/L	1.00	1.00	1.0

DATA REPORTING QUALIFIERS

Client: Sullivan County BOCES

Job Number:

Sdg Number: Monticello St John St High/Middle School

Lab Section	Qualifier	Description
Metals		
	g	Result fails applicable NYS drinking water standards
	U	The analyte was analyzed for but not detected at or above the lowest stated limit.

Certification Information

Client: Sullivan County BOCES

Job Number:

Sdg Number: Monticello St John St High/Middle School

The following analytes are Not Part of the ELAP scope of accreditation:

Sulfur, Tungsten, Silicon, Bicarbonate Alkalinity, 7 Day BOD 5210C, 28 Day BOD, Soluble BOD, Carbon Dioxide, Carbonate Alkalinity, CBOD Soluble, Chlorine, Cyanide (WAD), Ferrous Iron, Ferric Iron, Total Nitrogen, Total Organic Nitrogen, Dissolved Oxygen, pH, Phenolphthalein Alkalinity, Solids (Fixed), Solids (Percent), Solids (Percent Moisture), Solids (Percent Volatile), Solids (Volatile Suspended), Temperature, TKN (Soluble), Total Inorganic Carbon, Volatile Acids as Acetic Acid, 2-Aminopyridine, 3-Picoline, 1-Methyl-2-pyrrilidinone, Aziridine, Dimethyl sulfoxide, 1-Chlorohexane, Iron Bacteria, Salmonella, & Sulfur Reducing Bacteria.

The following analytes are Not Part of ELAP Potable Water scope of accreditation:

Cobalt (200.7, 200.8), Tin (200.7), Strontium (200.7), Gold (200.7), Platinum (200.7), Palladium (200.7), Titanium (200.7), Phosphorus (365.3), Nitrate-Nitrite (10-107-4-1C, 353.2), m-Xylene & p-Xylene (502.2, 524), Naphthalene (502.2), o-Xylene (502.2, 524), & Fecal Coliform (9222D).

The following analytes are Not Part of ELAP Solid and Hazardous Waste scope of accreditation:

Ammonia (SM 4500NH3G), TKN (351.2), Phosphorus (365.3), 1,2-Dichloro-1,1,2-trifluoroethane (8260), & Chlorodifluoromethane (8260).

The following analytes are Not Part of ELAP Non Potable Water scope of accreditation:

Dissolved Organic Carbon (5310C), Mecoprop (8151A), & MCPA (8151A).

Definitions and Glossary

Client: Sullivan County BOCES

Job Number:

Sdg Number: Monticello St John St High/Middle School

<u>Abbreviation</u>	<u>These commonly used abbreviations may or may not be present in this report.</u>
%R	Percent Recovery
DL, RA, RE	Indicates a Dilution, Reanalysis or Reextraction.
EPA	United States Environmental Protection Agency
MDL	Method Detection Limit - an estimate of the minimum amount of a substance that an analytical process can reliably detect. A MDL is analyte- and matrix-specific and may be laboratory-dependent.
ND	Not detected at the reporting limit (or MDL if shown).
QC	Quality Control
RL	Reporting Limit - the minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.
RPD	Relative Percent Difference - a measure of the relative difference between two points.

MONTICELLO- ST JOHN ST/ HIGH/ MIDDLE SCHOOL OUTSIDE FAUCETS

111923

water samples

Sample #	Date	Time	AREA	250 ml HNO3	Analysis Requested
001	10/18/16	7AM	LADIES BATH NEAR SECURITY OFFICE- LEFT SINK		LEAD
002			LADIES BATH NEAR SECURITY OFFICE- RIGHT SINK		
003			TEACHERS LOUNGE BATH SINK- RM 305		
004			HEALTH OCC LAB #1- RM 307		
005			HEALTH OCC LAB #2- RM 309		
006			TEACHERS LOUNGE BATH SINK - RM 227		
007			RM 127 SLOP SINK		
008			RM 125 OFFICE BATH SINK		
009			RM 125 CLASSROOM BATH SINK- LEFT		
010			RM 125 CLASSROOM BATH SINK - RIGHT		
011			RM 124A BATH SINK		
012			RM 124 KITCHEN BATH SINK		
013			RM 123 BATH SINK		
014			STAIRCASE "B" BATHROOM SINK		
015			RM 116 BATH SINK		
016			RM 115 BATH SINK		
017			BASEMENT WORKSHOP SLOP SINK		
018			CONSTRUCTION CLASSROOM SINK- LEFT		
019			CONSTRUCTIO CLASSROOM SINK- RIGHT		
020			HIGH SCHOOL OUTSIDE FAUCET		
021			HIGH SCHOOL OUTSIDE FAUCET		
022			HIGH SCHOOL OUTSIDE FAUCET		
023			HIGH SCHOOL OUTSIDE FAUCET		
024			MIDDLE SCHOOL OUTSIDE FAUCET		
025			HIGH SCHOOL OUTSIDE FAUCET		
026			HIGH SCHOOL OUTSIDE FAUCET		

Sampled By: *[Signature]*
 Relinquished By: *[Signature]*

Received by: *[Signature]* ER
 10/18/16 1545 25.5°C
 NO ICE
 IR#3

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Sullivan County BOCES

Job Number: 420-111923-1

SDG Number: Monticello St John St High/Middle School

Login Number: 111923

Question	T/F/NA	Comment
Samples were collected by ETL employee as per SOP-SAM-1	NA	
The cooler's custody seal, if present, is intact.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is recorded.	True	25.5 C
Cooler Temp. is within method specified range.(0-6 C PW, 0-8 C NPW, or BAC <10 C	True	
If false, was sample received on ice within 6 hours of collection.	NA	
Based on above criteria cooler temperature is acceptable.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	NA	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	